

A sunset over a body of water with mountains in the distance. The sky is a mix of blue, orange, and yellow, with the sun low on the horizon. The water is calm and reflects the colors of the sky. The mountains are dark and silhouetted against the horizon.


QUANTITATIVE NMR IN PROFILING OF BIOREFINERY PRODUCTS

Reino Laatikainen,
Department of Pharmacy, UEF,
Kuopio, Finland

<http://chemadder.com>

NMR METABONOMICS LABORATORY (MLAB)

HIGH-THROUGHPUT NMR METABONOMICS




- Sample into magnet
- Heat sample to +37°C
- Tune & Homogenise magnetic field
- Measure data
- Analyze data
- Make conclusions


>200 000 Serum samples (>600 000 spectra) in 2009-2014 !

QMTLS (Quantum Mechanical Total-Line-Shape) Analysis


User interface of program qQMTLS




NMR Spectrum (at 600 MHz) of an ABOWE sample



Aliphatic Region




2,3-Butanediol has a unique signal




A compound can be identified also from splittings (coupling constants) of multiplets: couplings do not depend on instrument or sample

Detection of mannitol



Spiking is sometimes used to ensure identification a component in complex samples

Quantum Mechanical Spectral Analysis (QMSA)



Chemical shift (ν) = weight point of multiplet


Coupling constant (J) \approx difference of two lines \Rightarrow fine structure

INDEPENDENT OF INSTRUMENTATION ..unlike MS!!

If chemical shifts, coupling constants & line-shape are given, spectrum (even the smallest details) can be simulated quantum mechanically !

ν_A, ν_B, ν_C
 J_{AB}, J_{AC}, J_{BC}

QMTLS



=> Model spectra (for quantitative analysis)

A problem: line-widths and chemical shifts (less) depend on sample, which means that a sophisticated software is needed for accurate quantification, ...a simple regression analysis does not work.

Adaptive Spectrum Libraries:

Analyze spectrum with one (magnetic) field, the spectrum at any other field can be then simulated !

A GOOD OBSERVED SPECTRUM
AT 600 MHz

400 MHz

600 MHz


800 MHz

Simulated


Original
Spectral
Analysis

Simulated

3.900 3.850 3.800 3.750 3.700



Quantitative QMSA of an ABOWE sample using 23 metabolites:



Sometimes spectral lines are broadened by Fe & Mn-ions, like above.


REPORT

&QM NAME	N	PROTONS	POPULATION	MOL%	mMOL	<u>Weight (mg/ml)</u>
%Q lactate	1	4	0.9004E+01	9.0918	86.3675	7.7731
%Q acetate	2	3	0.1680E+02	16.9621	161.1321	9.6679
%Q ala	3	4	0.6009E+01	6.0681	57.6442	5.1303
%Q valine	4	8	0.2383E+01	2.4059	22.8546	2.5140
%Q leu	5	10	0.2505E+01	2.5291	24.0256	3.1474
%Q ile	6	10	0.1930E+01	1.9491	18.5155	2.4255
%Q etoh	7	5	0.7797E+01	7.8735	74.7941	3.4405
%Q butyrate	8	7	0.7870E+00	0.7947	7.5494	0.6643
%Q propio	9	5	0.1874E+01	1.8922	17.9750	1.3661
%Q glu	10	5	0.9932E-02	0.0100	0.0953	0.0141
%Q beta	11	7	0.8427E+01	8.5091	80.8322	14.5498
%Q alfa	12	7	0.5282E+01	5.3338	50.6687	9.1204
%Q gly	13	2	0.1311E+01	1.3236	12.5734	0.9430
%Q thr	14	5	0.9269E+00	0.9360	8.8912	1.0581
%Q phe	15	8	0.1850E+01	1.8678	17.7434	2.6793
%Q 3pheprop	16	9	0.2500E+00	0.2524	2.3980	0.3597
%Q creatine	17	5	0.1144E+01	1.1552	10.9735	0.9766
%Q gaba	18	6	0.9932E-02	0.0100	0.0953	0.0098
%Q asp	19	3	0.1473E+01	1.4870	14.1258	1.8787
%Q mannitol	20	8	0.9783E+01	9.8790	93.8454	17.0799
%Q 23bud	21	8	0.1504E+02	15.1904	144.3013	12.9871
%Q sucrose	22	14	0.4436E+01	4.4791	42.5497	16.0838
%Q tsp	23	9	0.9685E+00	0.9779	9.2900	1.3573
TOTAL(excl. reference) =			99.0315	100.0000	949.9510	<u>113.8694</u>


QMTLS - APPLICATIONS

- Up to 100 metabolites in one sample?
- Dynamic range of 0.01-100 mol%
- Concentrations > 0.01 -M
- Applications:
 - Any mixtures and impurity analysis
 - Biofluids: plasma, CSF, lipid extracts of serum, urine,
 - Bioextracts, juices, ...

Linearity & confidence limits



Standard deviation vs. mol% %



Calculated vs. real impurity concentrations (in mol%)
 $R^2 = 0.995$

NO CALIBRATION NEEDED !!

CONS & PROS

- + Sample preparation, ..just filter and add reference
- + No calibration
- + Semiquantitative analysis of sample at one glance
- + Chemical confidence (identification of components directly from spectrum),..also carbohydrates (not with MS)
- + ASL (Adaptive Spectrum Libraries)
- + Almost automatical analysis

- Some expertise needed
- Not very sensitive, sample size > 0.3 ml
- Instrumentation (ca. 20€/sample, depends on n), ..liquid Helium and Nitrogen

COCLUSIONS and answers to questions presented after speech

Is ON-LINE possible with NMR ?

- In principle, yes, in fact NMR could allow automatic follow-up of the process once in a few minutes.
- Unfortunately, not yet feasible with the presently available instruments, ..but probably in near future with the new instruments (previous slide).

Are the new low field (40-100 MHz) instruments useful?

- Not for the water solutions! The minimum useful field is probably 200-400 MHz and demands far better water suppression and sensitivity than in the new < 100 MHz instruments.

RECOMMENDATION:

- NMR is invaluable in checking composition and detecting metabolites (especially sugars) of fermentation products, whenever starting materials or protocols are changed.
- NMR suits perfectly to calibration of methods like GC and HPLC; it is not necessary to prepare the calibration samples containing accurate known concentrations of metabolites (which may be unavailable).